



Corporate Accreditation No 63
Accredited for compliance with ISO/IEC 17025 - Testing



Analytical Report 309961

Issue Date: 14/08/2024
Issued By : Sydney Water Laboratory Services

Delivery Address: Sydney Water Corporation
51 Hermitage Rd
West Ryde NSW 2114
Telephone: (02) 9800 6935
Email: analyticalservices@sydneywater.com.au

Attention: Keira Anderson-King
Customer: Mid-Coast Council
Customer ID: ZFEL

Address: PO Box 482, Taree NSW 2430
Telephone: (02) 7955 7805
Email: [Keira Anderson-King@MidCoast.nsw.gov.au](mailto:Keira.Anderson-King@MidCoast.nsw.gov.au)

CONTENTS

1. Sydney Water Approved Signatory
2. Sample Summary
3. Analytical results
4. Comments
5. Laboratory QC results

Sydney Water Approved Signatory

Dwijal Dave, Clean & Waste Water Analyst

Abhilasha Bhandari, Organics Analyst

Justice Esoun-Nyarkoh, Organics Analyst

Where a result is required to meet a compliance limit or specification the associated uncertainty must be considered. Uncertainty estimates are available for all accredited test results.

SAMPLE SUMMARY

<u>Client Sample ID</u>	<u>Sample Number</u>	<u>Sampling Procedure</u>	<u>Date Sampled</u>	<u>Date Received</u>	<u>Date Authorised</u>	<u>Description</u>
M241303/1	L24066779	1	31/07/2024	02/08/2024	13/08/2024	MB:01 (EW)
M241303/2	L24066780	1	31/07/2024	02/08/2024	13/08/2024	MB:02 (EW)
M241303/3	L24066781	1	31/07/2024	02/08/2024	13/08/2024	LP:01 (EW)

Sampling procedures

- 1 Samples analysed as received.
- 2 Samples collected as per FS procedures SAWI 070, Excluding Oil & Grease which is collected as per clients instructions.
- 3 Samples collected as per FS procedures SAWI 070.
- 4 Results reported as received from WNSW.

ANALYTICAL RESULTS

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
CHEMISTRY									
WC26NS250 : Oil & Grease									
Oil & Grease	mg/L	<10	<10	<10					
Pet H/C	mg/L	<10	<10	<10					
Date of Performance	DD/MM/YY	06/08/24	06/08/24	06/08/24					
EXTERNAL TESTING									
XAL_PHNT : Total Phenolics analysed by ALS									
Courier*	N/A	ATC	ATC	ATC					
Total Phenolics	mg/L	<0.05	<0.05	<0.05					
XEN_CR6 : Hexavalent Chromium by Envirolab									
Hexavalent Chromium	mg/L	<0.001	<0.001	<0.001					
ORGANICS									
TC001WLL : Organochlorine Pesticides									

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC001WLL : Organochlorine Pesticides(Continued)									
Hexachlorobenzene	ug/L	<0.01	<0.01	<0.01					
alpha-BHC	ug/L	<0.01	<0.01	<0.01					
Lindane (gamma-BHC)	ug/L	<0.01	<0.01	<0.01					
Heptachlor	ug/L	<0.01	<0.01	<0.01					
Aldrin	ug/L	<0.01	<0.01	<0.01					
beta-BHC	ug/L	<0.01	<0.01	<0.01					
delta-BHC	ug/L	<0.01	<0.01	<0.01					
Heptachlor Epoxide	ug/L	<0.01	<0.01	<0.01					
alpha-Endosulfan	ug/L	<0.01	<0.01	<0.01					
gamma-Chlordane	ug/L	<0.01	<0.01	<0.01					
alpha-Chlordane	ug/L	<0.01	<0.01	<0.01					
4,4-DDE	ug/L	<0.01	<0.01	<0.01					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC001WLL : Organochlorine Pesticides(Continued)									
Dieldrin	ug/L	<0.01	<0.01	<0.01					
Endrin	ug/L	<0.01	<0.01	<0.01					
4,4-DDD	ug/L	<0.01	<0.01	<0.01					
beta-Endosulfan	ug/L	<0.01	<0.01	<0.01					
4,4-DDT	ug/L	<0.01	<0.01	<0.01					
Methoxychlor	ug/L	<0.01	<0.01	<0.01					
Endosulfan Sulphate	ug/L	<0.01	<0.01	<0.01					
Total Chlordane	ug/L	<0.01	<0.01	<0.01					
OC surrogate*	% Recovery	95	87	80					
Aldrin + Dieldrin*	ug/L	<0.01	<0.01	<0.01					
Endrin aldehyde*	ug/L	<0.01	<0.01	<0.01					
Endrin ketone*	ug/L	<0.01	<0.01	<0.01					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC001WLL : Organochlorine Pesticides(Continued)									
Total Tested DDT*	ug/L	<0.01	<0.01	<0.01					
Date of Performance	DD/MM/YY	06/08/24	06/08/24	06/08/24					
TC003WLL : Polychlorinated Biphenyls (PCBs)									
PCB TOTAL	ug/L	<0.1	<0.1	<0.1					
PCB surrogate*	% Recovery	95	87	80					
Date of Performance	DD/MM/YY	06/08/24	06/08/24	06/08/24					
TC004WLL : Polynuclear Aromatic Hydrocarbons									
Naphthalene	ug/L	<0.1	<0.1	<0.1					
Acenaphthene	ug/L	<0.1	<0.1	<0.1					
Fluorene	ug/L	<0.1	<0.1	<0.1					
Phenanthrene	ug/L	<0.1	<0.1	<0.1					
Anthracene	ug/L	<0.1	<0.1	<0.1					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC004WLL : Polynuclear Aromatic Hydrocarbons(Continued)									
Fluoranthene	ug/L	<0.1	<0.1	<0.1					
Pyrene	ug/L	<0.1	<0.1	<0.1					
Benzo(a)anthracene	ug/L	<0.1	<0.1	<0.1					
Chrysene	ug/L	<0.1	<0.1	<0.1					
Benzo(b)fluoranthene	ug/L	<0.1	<0.1	<0.1					
Benzo(k)fluoranthene	ug/L	<0.1	<0.1	<0.1					
Benzo(e)pyrene	ug/L	<0.1	<0.1	<0.1					
Benzo(a)pyrene	ug/L	<0.1	<0.1	<0.1					
Indeno(1,2,3-cd)pyrene	ug/L	<0.1	<0.1	<0.1					
Dibenzo(a,h)anthracene	ug/L	<0.1	<0.1	<0.1					
Benzo(ghi)perylene	ug/L	<0.1	<0.1	<0.1					
Total Detectable TDPAH	ug/L	<0.1	<0.1	<0.1					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC004WLL : Polynuclear Aromatic Hydrocarbons(Continued)									
PAH surrogate*	% Recovery	84	87	84					
Benzo(a)pyrene TEQ	ug/L	<0.1	<0.1	<0.1					
Acenaphthylene	ug/L	<0.1	<0.1	<0.1					
Perylene	ug/L	<0.1	<0.1	<0.1					
Date of Performance	DD/MM/YY	06/08/24	06/08/24	06/08/24					
TC0056GW : Multi-Residue Pesticide Analysis by UPLCMSMS									
Dichlorvos	ug/L	<0.05	<0.05	<0.05					
Mevinphos	ug/L	<0.05	<0.05	<0.05					
Ethoprop	ug/L	<0.05	<0.05	<0.05					
Demeton-S-Methyl	ug/L	<0.05	<0.05	<0.05					
Diazinon	ug/L	<0.05	<0.05	<0.05					
Fonofos	ug/L	<0.05	<0.05	<0.05					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC0056GW : Multi-Residue Pesticide Analysis by UPLCMSMS(Continued)									
Isazophos	ug/L	<0.05	<0.05	<0.05					
Dimethoate	ug/L	<0.05	<0.05	<0.05					
Methyl chlorpyrifos	ug/L	<0.05	<0.05	<0.05					
Ethyl chlorpyrifos	ug/L	<0.05	<0.05	<0.05					
Malathion	ug/L	<0.05	<0.05	<0.05					
Fenthion	ug/L	<0.05	<0.05	<0.05					
Fenitrothion	ug/L	<0.1	<0.1	<0.1					
Parathion	ug/L	<0.05	<0.05	<0.05					
Phospholan	ug/L	<0.05	<0.05	<0.05					
Fensulfothion	ug/L	<0.05	<0.05	<0.05					
E.P.N	ug/L	<0.1	<0.1	<0.1					
Methyl azinphos	ug/L	<0.05	<0.05	<0.05					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC0056GW : Multi-Residue Pesticide Analysis by UPLCMSMS(Continued)									
Ethyl azinphos	ug/L	<0.1	<0.1	<0.1					
metsulfuron-methyl	ug/L	<0.1	<0.1	<0.1					
Molinate	ug/L	<0.05	<0.05	<0.05					
Atrazine	ug/L	<0.05	<0.05	<0.05					
Chlorfenvinphos	ug/L	<0.05	<0.05	<0.05					
Pirimiphos methyl	ug/L	<0.05	<0.05	<0.05					
Propiconazole	ug/L	<0.05	<0.05	<0.05					
Simazine	ug/L	<0.05	<0.05	<0.05					
Bromophos ethyl	ug/L	<0.05	<0.05	<0.05					
Coumaphos	ug/L	<0.05	<0.05	<0.05					
Ethion	ug/L	<0.05	<0.05	<0.05					
Fenamiphos	ug/L	<0.05	<0.05	<0.05					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC0056GW : Multi-Residue Pesticide Analysis by UPLCMSMS(Continued)									
Monocrotophos	ug/L	<0.05	<0.05	<0.05					
Parathion Methyl	ug/L	<0.05	<0.05	<0.05					
Prothiofos	ug/L	<0.05	<0.05	<0.05					
Sulfotep	ug/L	<0.05	<0.05	<0.05					
Pirimiphos Ethyl	ug/L	<0.05	<0.05	<0.05					
Tetrachlorvinphos	ug/L	<0.05	<0.05	<0.05					
Propazine	ug/L	<0.05	<0.05	<0.05					
Trithion	ug/L	<0.1	<0.1	<0.1					
Diuron	ug/L	<0.5	<0.5	<0.5					
Hexazinone	ug/L	<0.1	0.1	0.5					
Temephos	ug/L	<1	<1	<1					
Carbaryl	ug/L	<0.1	<0.1	<0.1					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC0056GW : Multi-Residue Pesticide Analysis by UPLCMSMS(Continued)									
Disulfoton	ug/L	<0.1	<0.1	<0.1					
Phorate	ug/L	<0.1	<0.1	<0.1					
Acephate	ug/L	<0.5	<0.5	<0.5					
Aldicarb	ug/L	<0.1	<0.1	<0.1					
OP surrogate*	% Recovery	101.10	91.54	83.60					
Carbofuran	ug/L	<0.05	<0.05	<0.05					
Methomyl	ug/L	<0.025	<0.025	<0.025					
organophosphate pesticides(Total)	ug/L	<2.5	<2.5	<2.5					
Date of Performance	DD/MM/YY	08/08/24	08/08/24	08/08/24					
TC010304PW : Sample Prep for Organics analysis									
OC/PCB Preparation in Liquid Samples	N/A	Done	Done	Done					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID	M241303/1	M241303/2	M241303/3					
Sampled Date	31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number	L24066779	L24066780	L24066781					
ORGANICS								
TC010304PW : Sample Prep for Organics analysis(Continued)								
PAH Preparation in Liquid Samples	N/A	Done	Done	Done				
Date of Performance	DD/MM/YY	06/08/24	06/08/24	06/08/24				
TC012WLL : Volatiles Organic Compounds								
Vinyl Chloride	ug/L	<0.3	<0.3	<0.3				
1,1-Dichloroethene	ug/L	<0.3	<0.3	<0.3				
Methylene Chloride	ug/L	1.7	0.5	1.9				
1,1-Dichloroethane	ug/L	<0.4	<0.4	<0.4				
Chloroform	ug/L	<0.5	<0.5	<0.5				
1,1,1-Trichloroethane	ug/L	<0.3	<0.3	<0.3				
Carbon Tetrachloride	ug/L	<0.3	<0.3	<0.3				
Benzene	ug/L	<0.3	<0.3	<0.3				
1,2-Dichloroethane	ug/L	<0.3	<0.3	<0.3				

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC012WLL : Volatiles Organic Compounds(Continued)									
Trichloroethene	ug/L	<0.3	<0.3	<0.3					
1,2-Dichloropropane	ug/L	<0.5	<0.5	<0.5					
Bromodichloromethane	ug/L	<0.5	<0.5	<0.5					
Toluene	ug/L	<0.3	<0.3	<0.3					
Tetrachloroethene	ug/L	<0.3	<0.3	<0.3					
Dibromochloromethane	ug/L	<0.5	<0.5	<0.5					
Chlorobenzene	ug/L	<0.3	<0.3	<0.3					
1,1,1,2-Tetrachloroethane	ug/L	<0.4	<0.4	<0.4					
Ethyl benzene	ug/L	<0.3	<0.3	<0.3					
(m+p)-Xylenes	ug/L	<0.4	<0.4	<0.4					
o-Xylene	ug/L	<0.4	<0.4	<0.4					
Styrene	ug/L	<0.4	<0.4	<0.4					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC012WLL : Volatiles Organic Compounds(Continued)									
Bromoform	ug/L	<0.5	<0.5	<0.5					
1,1,2,2-Tetrachloroethane	ug/L	<0.7	<0.7	<0.7					
m-Dichlorobenzene	ug/L	<0.3	<0.3	<0.3					
p-Dichlorobenzene	ug/L	<0.5	<0.5	<0.5					
1,1,2-Trichloroethane	ug/L	<0.3	<0.3	<0.3					
o-Dichlorobenzene	ug/L	<0.3	<0.3	<0.3					
cis 1,2 Dichloroethene	ug/L	<0.4	<0.4	<0.4					
trans 1,2 Dichloroethene	ug/L	<0.3	<0.3	<0.3					
Total 1,2-Dicloroethene	ug/L	<0.3	<0.3	<0.3					
1,2,3-Trichlorobenzene	ug/L	<0.3	<0.3	<0.3					
Volatiles surrogate*	% Recovery	100	110	99					
1,2,4-trimethylbenzene	ug/L	<0.5	<0.5	<0.5					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC012WLL : Volatiles Organic Compounds(Continued)									
1,2,4-Trichlorobenzene	ug/L	<0.4	<0.4	<0.4					
2-Chlorotoluene	ug/L	<0.4	<0.4	<0.4					
4-Chlorotoluene	ug/L	<0.4	<0.4	<0.4					
Bromobenzene	ug/L	<0.5	<0.5	<0.5					
Hexachlorobutadiene	ug/L	<0.3	<0.3	<0.3					
isopropylbenzene	ug/L	<0.3	<0.3	<0.3					
n-Butylbenzene	ug/L	<0.3	<0.3	<0.3					
n-propylbenzene	ug/L	<0.3	<0.3	<0.3					
p-isopropyltoluene	ug/L	<0.3	<0.3	<0.3					
sec-Butylbenzene	ug/L	<0.3	<0.3	<0.3					
tert-butylbenzene	ug/L	<0.3	<0.3	<0.3					
1,1-Dichloropropane	ug/L	<0.3	<0.3	<0.3					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC012WLL : Volatiles Organic Compounds(Continued)									
1,2,3-Trichloropropane	ug/L	<0.5	<0.5	<0.5					
1,3-Dichloropropane	ug/L	<0.5	<0.5	<0.5					
Bromochloromethane	ug/L	<0.4	<0.4	<0.4					
Bromomethane	ug/L	<0.4	<0.4	<0.4					
Chloroethane	ug/L	<0.3	<0.3	<0.3					
Chloromethane	ug/L	<0.4	<0.4	<0.4					
Dibromomethane	ug/L	<0.3	<0.3	<0.3					
Naphthalene	ug/L	<0.3	<0.3	<0.3					
Trichlorofluoromethane	ug/L	<0.3	<0.3	<0.3					
Dichlorodifluoromethane	ug/L	<0.3	<0.3	<0.3					
1,2-Dibromo-3-chloropropane	ug/L	<0.3	<0.3	<0.3					
2,2-Dichloropropane	ug/L	<0.2	<0.2	<0.2					

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		M241303/1	M241303/2	M241303/3					
Sampled Date		31/07/2024 08:30:00 AM	31/07/2024 09:25:00 AM	31/07/2024 10:00:00 AM					
Sample Number		L24066779	L24066780	L24066781					
ORGANICS									
TC012WLL : Volatiles Organic Compounds(Continued)									
cis-1,3-Dichloropropene	ug/L	<0.5	<0.5	<0.5					
trans-1,3-Dichloropropene	ug/L	<0.5	<0.5	<0.5					
1,2-dibromoethane	ug/L	<0.3	<0.3	<0.3					
1,3,5-trimethylbenzene	ug/L	<0.3	<0.3	<0.3					
1,3,5-Trichlorobenzene*	ug/L	<0.3	<0.3	<0.3					
Total Trihalomethanes	ug/L	<0.5	<0.5	<0.5					
Total tested Trichlorobenzenes*	ug/L	<0.4	<0.4	<0.4					
Total tested Xylenes*	ug/L	<0.4	<0.4	<0.4					
Date of Performance	DD/MM/YY	07/08/24	07/08/24	07/08/24					

COMMENTS

<u>Sample ID</u>	<u>Comment Level</u>	<u>Method</u>	<u>Test</u>	<u>Comment</u>
L24066779	Method	XAL_PHNT	-	Sample analysed by ALS WO #ES2425392
L24066780	Method	XAL_PHNT	-	Sample analysed by ALS WO #ES2425392

* Indicates NATA accreditation does not cover the performance of this service

L24066781	Method	XAL_PHNT	-	Sample analysed by ALS WO #ES2425392
L24066779	Method	XEN_CR6	-	Sample analysed by Envirolab Ref # 358325
L24066780	Method	XEN_CR6	-	Sample analysed by Envirolab Ref # 358325
L24066781	Method	XEN_CR6	-	Sample analysed by Envirolab Ref # 358325

Analysis

XEN_CR6
XAL_PHNT

Analysis Requirements

NATA accreditation number 2901
Analysed by ALS NATA Accreditation No.825

* Indicates NATA accreditation does not cover the performance of this service

LABORATORY QC RESULTS

N/A - Not Applicable

PQL - Practical Quantitation Limit

LOQ - Limit of Quantification

RPD - Relative Percent Difference

SPIKE/Positive Control - Addition of a known amount and concentration

Duplicate Precision = Accepted - Result 2 within 95% confidence limits of result 1

Duplicate Precision = Outlier - Result 2 outside 95% confidence limits of result 1

Duplicate Precision = Not calculated - Result is outside test range

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC001WLL 4,4-DDD						
<0.01 ug/L	<0.01	107 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL 4,4-DDE						
<0.01 ug/L	<0.01	93.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL 4,4-DDT						
<0.01 ug/L	<0.01	101 70.0 - 130.0 ug/L	84 % Recovery 50.0 - 130.0 % Recovery	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Aldrin						
<0.01 ug/L	<0.01	98.0 70.0 - 130.0 ug/L	93 % Recovery 50.0 - 130.0 % Recovery	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL alpha-BHC						
<0.01 ug/L	<0.01	98.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL alpha-Chlordane						
<0.01 ug/L	<0.01	95.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC001WLL alpha-Endosulfan						
<0.01 ug/L	<0.01	93.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL beta-BHC						
<0.01 ug/L	<0.01	98.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL beta-Endosulfan						
<0.01 ug/L	<0.01	105 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL delta-BHC						
<0.01 ug/L	<0.01	108 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Dieldrin						
<0.01 ug/L	<0.01	93.0 70.0 - 130.0 ug/L	90 % Recovery 50.0 - 130.0 % Recovery	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Endosulfan Sulphate						
<0.01 ug/L	<0.01	102 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Endrin						
<0.01 ug/L	<0.01	96.0 70.0 - 130.0 ug/L	86 % Recovery 50.0 - 130.0 % Recovery	<0.01	<0.01	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC001WLL gamma-Chlordane						
<0.01 ug/L	<0.01	94.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Heptachlor						
<0.01 ug/L	<0.01	96.0 70.0 - 130.0 ug/L	87 % Recovery 50.0 - 130.0 % Recovery	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Heptachlor Epoxide						
<0.01 ug/L	<0.01	101 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Hexachlorobenzene						
<0.01 ug/L	<0.01	101 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Lindane (gamma-BHC)						
<0.01 ug/L	<0.01	97.0 70.0 - 130.0 ug/L	89 % Recovery 50.0 - 130.0 % Recovery	<0.01	<0.01	B 0.0 - 0.0 %
TC001WLL Methoxychlor						
<0.01 ug/L	<0.01	95.0 70.0 - 130.0 ug/L	E	<0.01	<0.01	B 0.0 - 0.0 %
TC003WLL PCB TOTAL						
<0.1 ug/L	<0.1	94.0 70.0 - 130.0 ug/L	98.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC004WLL Acenaphthene						
<0.1 ug/L	<0.1	99.0 <i>70.0 - 130.0 ug/L</i>	90.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Acenaphthylene						
<0.1 ug/L	<0.1	92.0 <i>70.0 - 130.0 ug/L</i>	82.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Anthracene						
<0.1 ug/L	<0.1	90.0 <i>70.0 - 130.0 ug/L</i>	80.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Benzo(a)anthracene						
<0.1 ug/L	<0.1	107 <i>70.0 - 130.0 ug/L</i>	114 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Benzo(a)pyrene						
<0.1 ug/L	<0.1	112 <i>70.0 - 130.0 ug/L</i>	118 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Benzo(b)fluoranthene						
<0.1 ug/L	<0.1	98.0 <i>70.0 - 130.0 ug/L</i>	115 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Benzo(e)pyrene						
<0.1 ug/L	<0.1		103 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC004WLL Benzo(ghi)perylene						
<0.1 ug/L	<0.1	95.0 <i>70.0 - 130.0 ug/L</i>	98.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Benzo(k)fluoranthene						
<0.1 ug/L	<0.1	92.0 <i>70.0 - 130.0 ug/L</i>	83.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Chrysene						
<0.1 ug/L	<0.1	97.0 <i>70.0 - 130.0 ug/L</i>	99.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Dibenzo(a,h)anthracene						
<0.1 ug/L	<0.1	85.0 <i>70.0 - 130.0 ug/L</i>	119 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Fluoranthene						
<0.1 ug/L	<0.1	122 <i>70.0 - 130.0 ug/L</i>	78.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Fluorene						
<0.1 ug/L	<0.1	96.0 <i>70.0 - 130.0 ug/L</i>	84.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC004WLL Indeno(1,2,3-cd)pyrene						
<0.1 ug/L	<0.1	85.0 <i>70.0 - 130.0 ug/L</i>	126 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC004WLL Naphthalene						
<0.1 ug/L	<0.1	99.0 70.0 - 130.0 ug/L	90.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC004WLL Perylene						
<0.1 ug/L	<0.1		96.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC004WLL Phenanthrene						
<0.1 ug/L	<0.1	97.0 70.0 - 130.0 ug/L	90.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC004WLL Pyrene						
<0.1 ug/L	<0.1	116 70.0 - 130.0 ug/L	78.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0056GW Acephate						
<0.5 ug/L	<0.5	114.0 70.0 - 130.0 ug/L	114.0 % Recovery 50.0 - 130.0 % Recovery	<0.5	<0.5	B 0.0 - 0.0 %
TC0056GW Aldicarb						
<0.1 ug/L	<0.1	120.0 70.0 - 130.0 ug/L	103.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0056GW Atrazine						
<0.05 ug/L	<0.05	115.0 70.0 - 130.0 ug/L	115.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Bromophos ethyl						
<0.05 ug/L	<0.05	117.0 <i>70.0 - 130.0 ug/L</i>	109.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Carbaryl						
<0.1 ug/L	<0.1	112.0 <i>70.0 - 130.0 ug/L</i>	110.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC0056GW Carbofuran						
<0.05 ug/L	<0.05	115.0 <i>70.0 - 130.0 ug/L</i>	114.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Chlorfenvinphos						
<0.05 ug/L	<0.05	121.0 <i>70.0 - 130.0 ug/L</i>	123.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Coumaphos						
<0.05 ug/L	<0.05	107.0 <i>70.0 - 130.0 ug/L</i>	120.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Demeton-S-Methyl						
<0.05 ug/L	<0.05	129.0 <i>70.0 - 130.0 ug/L</i>	116.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Diazinon						
<0.05 ug/L	<0.05	118.0 <i>70.0 - 130.0 ug/L</i>	119.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Dichlorvos						
<0.05 ug/L	<0.05	109 70.0 - 130.0 ug/L	104 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Dimethoate						
<0.05 ug/L	<0.05	98.8 70.0 - 130.0 ug/L	99.3 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Disulfoton						
<0.1 ug/L	<0.1	81.3 70.0 - 130.0 ug/L	82.5 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0056GW Diuron						
<0.5 ug/L	<0.5	113.0 70.0 - 130.0 ug/L	113.0 % Recovery 50.0 - 130.0 % Recovery	<0.5	<0.5	B 0.0 - 0.0 %
TC0056GW E.P.N						
<0.1 ug/L	<0.1	115.0 70.0 - 130.0 ug/L	110.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0056GW Ethion						
<0.05 ug/L	<0.05	122.0 70.0 - 130.0 ug/L	117.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Ethoprop						
<0.05 ug/L	<0.05	105.0 70.0 - 130.0 ug/L	115.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Ethyl azinphos						
<0.1 ug/L	<0.1	122.0 <i>70.0 - 130.0 ug/L</i>	73.1 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC0056GW Ethyl chlorpyrifos						
<0.05 ug/L	<0.05	124.0 <i>70.0 - 130.0 ug/L</i>	126.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Fenamiphos						
<0.05 ug/L	<0.05	108.0 <i>70.0 - 130.0 ug/L</i>	120.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Fenitrothion						
<0.1 ug/L	<0.1	114.0 <i>70.0 - 130.0 ug/L</i>	114.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC0056GW Fensulfothion						
<0.05 ug/L	<0.05	98.5 <i>70.0 - 130.0 ug/L</i>	104.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Fenthion						
<0.05 ug/L	<0.05	88.3 <i>70.0 - 130.0 ug/L</i>	87.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Fonofos						
<0.05 ug/L	<0.05	116.0 <i>70.0 - 130.0 ug/L</i>	113.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Hexazinone						
<0.1 ug/L	<0.1	113.0 70.0 - 130.0 ug/L	121.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0056GW Isazophos						
<0.05 ug/L	<0.05	126.0 70.0 - 130.0 ug/L	119.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Malathion						
<0.05 ug/L	<0.05	114.0 70.0 - 130.0 ug/L	112.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Methomyl						
<0.025 ug/L	<0.025	114.0 70.0 - 130.0 ug/L	116.0 % Recovery 50.0 - 130.0 % Recovery	<0.025	<0.025	B 0.0 - 0.0 %
TC0056GW Methyl azinphos						
<0.05 ug/L	<0.05	104.0 70.0 - 130.0 ug/L	107.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Methyl chlorpyrifos						
<0.05 ug/L	<0.05	117.0 70.0 - 130.0 ug/L	124.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW metsulfuron-methyl						
<0.1 ug/L	<0.1	128.0 70.0 - 130.0 ug/L	122.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Mevinphos						
<0.05 ug/L	<0.05	77.4 70.0 - 130.0 ug/L	78.9 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Molinate						
<0.05 ug/L	<0.05	111.0 70.0 - 130.0 ug/L	109.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Monocrotophos						
<0.05 ug/L	<0.05	114.0 70.0 - 130.0 ug/L	114.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Parathion						
<0.05 ug/L	<0.05	108.0 70.0 - 130.0 ug/L	109.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Parathion Methyl						
<0.05 ug/L	<0.05	115.0 70.0 - 130.0 ug/L	116.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0056GW Phorate						
<0.1 ug/L	<0.1	71.7 70.0 - 130.0 ug/L	85.0 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0056GW Phospholan						
<0.05 ug/L	<0.05	120.0 70.0 - 130.0 ug/L	120.0 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Pirimiphos Ethyl						
<0.05 ug/L	<0.05	127.0 <i>70.0 - 130.0 ug/L</i>	121.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Pirimiphos methyl						
<0.05 ug/L	<0.05	118.0 <i>70.0 - 130.0 ug/L</i>	125.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Propazine						
<0.05 ug/L	<0.05	104.0 <i>70.0 - 130.0 ug/L</i>	103.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Propiconazole						
<0.05 ug/L	<0.05	110.0 <i>70.0 - 130.0 ug/L</i>	117.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Prothiofos						
<0.05 ug/L	<0.05	126.0 <i>70.0 - 130.0 ug/L</i>	101.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Simazine						
<0.05 ug/L	<0.05	110.0 <i>70.0 - 130.0 ug/L</i>	106.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Sulfotep						
<0.05 ug/L	<0.05	118.0 <i>70.0 - 130.0 ug/L</i>	117.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0056GW Temephos						
<1 ug/L	<1	123.0 <i>70.0 - 130.0 ug/L</i>	80.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<1	<1	B <i>0.0 - 0.0 %</i>
TC0056GW Tetrachlorvinphos						
<0.05 ug/L	<0.05	120.0 <i>70.0 - 130.0 ug/L</i>	118.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	B <i>0.0 - 0.0 %</i>
TC0056GW Trithion						
<0.1 ug/L	<0.1	114.0 <i>70.0 - 130.0 ug/L</i>	115.0 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.1	<0.1	B <i>0.0 - 0.0 %</i>
TC012WLL (m+p)-Xylenes						
<0.4 ug/L	<0.4	103 <i>70.0 - 130.0 ug/L</i>	114 % Recovery <i>50.0 - 130.0 % Recovery</i>	<1	<1	B <i>0.0 - 0.0 %</i>
TC012WLL 1,1,1,2-Tetrachloroethane						
<0.4 ug/L	<0.4		E	<2	<2	B <i>0.0 - 0.0 %</i>
TC012WLL 1,1,1-Trichloroethane						
<0.3 ug/L	<0.3		E	<3	<3	B <i>0.0 - 0.0 %</i>
TC012WLL 1,1,2,2-Tetrachloroethane						
<0.7 ug/L	<0.7		E	<2	<2	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL 1,1,2-Trichloroethane						
<0.3 ug/L	<0.3		E	<2	<2	B 0.0 - 0.0 %
TC012WLL 1,1-Dichloroethane						
<0.4 ug/L	<0.4		E	<2	<2	B 0.0 - 0.0 %
TC012WLL 1,1-Dichloroethene						
<0.3 ug/L	<0.3		E	<3	<3	B 0.0 - 0.0 %
TC012WLL 1,1-Dichloropropane						
<0.3 ug/L	F	C	E			
TC012WLL 1,2,3-Trichlorobenzene						
<0.3 ug/L	<0.3		E			
TC012WLL 1,2,3-Trichloropropane						
<0.5 ug/L	F	C	E			
TC012WLL 1,2,4-Trichlorobenzene						
<0.4 ug/L	<0.4		E			

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL 1,2,4-trimethylbenzene						
<0.5 ug/L	F	C	E			
TC012WLL 1,2-Dibromo-3-chloropropane						
<0.3 ug/L	F	C	E			
TC012WLL 1,2-dibromoethane						
<0.3 ug/L	F	C	E			
TC012WLL 1,2-Dichloroethane						
<0.3 ug/L	<0.3		E	<2	<2	B 0.0 - 0.0 %
TC012WLL 1,2-Dichloropropane						
<0.5 ug/L	<0.5		E	<2	<2	B 0.0 - 0.0 %
TC012WLL 1,3,5-Trichlorobenzene						
<0.3 ug/L	F	C	E			
TC012WLL 1,3,5-trimethylbenzene						
<0.3 ug/L	F	C	E			

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL 1,3-Dichloropropane						
<0.5 ug/L	F	C	E			
TC012WLL 2,2-Dichloropropane						
<0.2 ug/L	F	C	E			
TC012WLL 2-Chlorotoluene						
<0.4 ug/L	<0.4		E			
TC012WLL 4-Chlorotoluene						
<0.4 ug/L	<0.4		E			
TC012WLL Benzene						
<0.3 ug/L	<0.3	102 70.0 - 130.0 ug/L	118 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL Bromobenzene						
<0.5 ug/L	<0.5		E			
TC012WLL Bromochloromethane						
<0.4 ug/L	F	C	E			

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL Bromodichloromethane						
<0.5 ug/L	<0.5	94 70.0 - 130.0 ug/L	112 % Recovery 50.0 - 130.0 % Recovery	3	3	B 0.0 - 0.0 %
TC012WLL Bromoform						
<0.5 ug/L	<0.5	96 70.0 - 130.0 ug/L	119 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL Bromomethane						
<0.4 ug/L	F	C	E			
TC012WLL Carbon Tetrachloride						
<0.3 ug/L	<0.3		E	<3	<3	B 0.0 - 0.0 %
TC012WLL Chlorobenzene						
<0.3 ug/L	<0.3	106 70.0 - 130.0 ug/L	124 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL Chloroethane						
<0.3 ug/L	F	C	E			
TC012WLL Chloroform						
<0.5 ug/L	<0.5	97 70.0 - 130.0 ug/L	114 % Recovery 50.0 - 130.0 % Recovery	17	18	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL Chloromethane						
<0.4 ug/L	F	C	E			
TC012WLL cis 1,2 Dichloroethene						
<0.4 ug/L	<0.4		E	<1	<1	B 0.0 - 0.0 %
TC012WLL cis-1,3-Dichloropropene						
<0.5 ug/L	F	C	E			
TC012WLL Dibromochloromethane						
<0.5 ug/L	<0.5	92 70.0 - 130.0 ug/L	115 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL Dichlorodifluoromethane						
<0.3 ug/L	F	C	E			
TC012WLL Ethyl benzene						
<0.3 ug/L	<0.3	107 70.0 - 130.0 ug/L	119 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL Hexachlorobutadiene						
<0.3 ug/L	F	C	E			

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL isopropylbenzene						
<0.3 ug/L	F	C	E			
TC012WLL m-Dichlorobenzene						
<0.3 ug/L	<0.3	98	114 % Recovery	<2	<2	B
		70.0 - 130.0 ug/L	50.0 - 130.0 % Recovery			0.0 - 0.0 %
TC012WLL Methylene Chloride						
<0.3 ug/L	<0.3		E	<3	<3	B
						0.0 - 0.0 %
TC012WLL Naphthalene						
<0.3 ug/L	F	C	E			
TC012WLL n-Butylbenzene						
<0.3 ug/L	<0.3		E			
TC012WLL n-propylbenzene						
<0.3 ug/L	F	C	E			
TC012WLL o-Dichlorobenzene						
<0.3 ug/L	<0.3	102	122 % Recovery	<2	<2	B
		70.0 - 130.0 ug/L	50.0 - 130.0 % Recovery			0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL o-Xylene						
<0.4 ug/L	<0.4	100 70.0 - 130.0 ug/L	112 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL p-Dichlorobenzene						
<0.5 ug/L	<0.5	90 70.0 - 130.0 ug/L	105 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL p-isopropyltoluene						
<0.3 ug/L	F	C	E			
TC012WLL sec-Butylbenzene						
<0.3 ug/L	<0.3		E			
TC012WLL Styrene						
<0.4 ug/L	<0.4	95 70.0 - 130.0 ug/L	105 % Recovery 50.0 - 130.0 % Recovery	<2	<2	B 0.0 - 0.0 %
TC012WLL tert-butylbenzene						
<0.3 ug/L	F	C	E			
TC012WLL Tetrachloroethene						
<0.3 ug/L	<0.3		E	<2	<2	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ		Blank	Control	Spike	Duplicate1	Duplicate2	RPD
			Acceptance Criteria	Acceptance Criteria	Acceptance Criteria		
TC012WLL Toluene							
<0.3 ug/L	<0.3	106 70.0 - 130.0 ug/L	118 % Recovery 50.0 - 130.0 % Recovery	<1	<1	B	
				0.0 - 0.0 %			
TC012WLL Total 1,2-Dicloroethene							
<0.3 ug/L	F	C	E	T			
TC012WLL Total tested Trichlorobenzenes							
<0.4 ug/L	F	C	E				
TC012WLL Total tested Xylenes							
<0.4 ug/L	F	C	E				
TC012WLL Total Trihalomethanes							
<0.5 ug/L	F	C	E				
TC012WLL trans 1,2 Dichloroethene							
<0.3 ug/L	<0.3		E	<3	<3	B	
				0.0 - 0.0 %			
TC012WLL trans-1,3-Dichloropropene							
<0.5 ug/L	F	C	E				

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC012WLL Trichloroethene						
<0.3 ug/L	<0.3		E	<3	<3	B 0.0 - 0.0 %
TC012WLL Trichlorofluoromethane						
<0.3 ug/L	F	C	E			
TC012WLL Vinyl Chloride						
<0.3 ug/L	<0.3		E	<3	<3	B 0.0 - 0.0 %
TC012WLL Volatiles surrogate						
	F	C	E	100	110	9.52 % 0.0 - 30.0 %
WC26NS250 Oil & Grease						
<10 mg/L	<10	88 85.1 - 101.5 mg/L	E			
WC26NS250 Pet H/C						
<3 mg/L	F		E			T

* Indicates NATA accreditation does not cover the performance of this service

Extra Note:

F: Blank is not applicable for this analyte

C: Standard/Control is not applicable for this analyte

T: Duplicate is not applicable for this analyte

E: Spike is not applicable for this analyte

DUPLICATE 4,4-DDD B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE 4,4-DDE B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE 4,4-DDT B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Aldrin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE alpha-BHC B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE alpha-Chlordane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE alpha-Endosulfan B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE beta-BHC B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE beta-Endosulfan B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE delta-BHC B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Dieldrin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Endosulfan Sulphate B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Endrin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE gamma-Chlordane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Heptachlor B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Heptachlor Epoxide B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Hexachlorobenzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Lindane (gamma-BHC) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Methoxychlor B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE PCB TOTAL B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Acenaphthene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Acenaphthylene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anthracene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Benzo(a)anthracene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Benzo(a)pyrene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Benzo(b)fluoranthene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Benzo(e)pyrene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Benzo(ghi)perylene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Benzo(k)fluoranthene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Chrysene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service

DUPLICATE Dibenzo(a,h)anthracene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fluoranthene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fluorene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Indeno(1,2,3-cd)pyrene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Naphthalene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Perylene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Phenanthrene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Pyrene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Acephate B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Aldicarb B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Atrazine B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Bromophos ethyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Carbaryl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Carbofuran B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Chlorfenvinphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Coumaphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Demeton-S-Methyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Diazinon B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Dichlorvos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Dimethoate B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Disulfoton B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Diuron B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE E.P.N B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Ethion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Ethoprop B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Ethyl azinphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Ethyl chlorpyrifos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fenamiphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fenitrothion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fensulfothion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fenthion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Fonofos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Hexazinone B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Isazophos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
DUPLICATE Malathion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service

DUPLICATE Methomyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Methyl azinphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Methyl chlorpyrifos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE metsulfuron-methyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Mevinphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Molinate B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Monocrotophos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Parathion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Parathion Methyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Phorate B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Phospholan B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Pirimiphos Ethyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Pirimiphos methyl B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Propazine B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Propiconazole B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Prothiofos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Simazine B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Sulfotep B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Temephos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Tetrachlorvinphos B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Trithion B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE (m+p)-Xylenes B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,1,1,2-Tetrachloroethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,1,1-Trichloroethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,1,2,2-Tetrachloroethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,1,2-Trichloroethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,1-Dichloroethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,1-Dichloroethene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD 1,1-Dichloropropane C: Standard/Control is not applicable for this analyte
 STANDARD 1,2,3-Trichloropropane C: Standard/Control is not applicable for this analyte
 STANDARD 1,2,4-trimethylbenzene C: Standard/Control is not applicable for this analyte
 STANDARD 1,2-Dibromo-3-chloropropane C: Standard/Control is not applicable for this analyte
 STANDARD 1,2-dibromoethane C: Standard/Control is not applicable for this analyte
 DUPLICATE 1,2-Dichloroethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE 1,2-Dichloropropane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service

STANDARD 1,3,5-Trichlorobenzene C: Standard/Control is not applicable for this analyte
 STANDARD 1,3,5-trimethylbenzene C: Standard/Control is not applicable for this analyte
 STANDARD 1,3-Dichloropropane C: Standard/Control is not applicable for this analyte
 STANDARD 2,2-Dichloropropane C: Standard/Control is not applicable for this analyte
 DUPLICATE Benzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Bromochloromethane C: Standard/Control is not applicable for this analyte
 DUPLICATE Bromodichloromethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Bromoform B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Bromomethane C: Standard/Control is not applicable for this analyte
 DUPLICATE Carbon Tetrachloride B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Chlorobenzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Chloroethane C: Standard/Control is not applicable for this analyte
 DUPLICATE Chloroform B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Chloromethane C: Standard/Control is not applicable for this analyte
 DUPLICATE cis 1,2 Dichloroethene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD cis-1,3-Dichloropropene C: Standard/Control is not applicable for this analyte
 DUPLICATE Dibromochloromethane B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Dichlorodifluoromethane C: Standard/Control is not applicable for this analyte
 DUPLICATE Ethyl benzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Hexachlorobutadiene C: Standard/Control is not applicable for this analyte
 STANDARD isopropylbenzene C: Standard/Control is not applicable for this analyte
 DUPLICATE m-Dichlorobenzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Methylene Chloride B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD Naphthalene C: Standard/Control is not applicable for this analyte
 STANDARD n-propylbenzene C: Standard/Control is not applicable for this analyte
 DUPLICATE o-Dichlorobenzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE o-Xylene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE p-Dichlorobenzene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD p-isopropyltoluene C: Standard/Control is not applicable for this analyte
 DUPLICATE Styrene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 STANDARD tert-butylbenzene C: Standard/Control is not applicable for this analyte
 DUPLICATE Tetrachloroethene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Toluene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
 DUPLICATE Total 1,2-Dichloroethene T: Duplicate is not applicable for this analyte
 STANDARD Total 1,2-Dichloroethene C: Standard/Control is not applicable for this analyte

* Indicates NATA accreditation does not cover the performance of this service

STANDARD Total tested Trichlorobenzenes C: Standard/Control is not applicable for this analyte
STANDARD Total tested Xylenes C: Standard/Control is not applicable for this analyte
STANDARD Total Trihalomethanes C: Standard/Control is not applicable for this analyte
DUPLICATE trans 1,2 Dichloroethene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
STANDARD trans-1,3-Dichloropropene C: Standard/Control is not applicable for this analyte
DUPLICATE Trichloroethene B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
STANDARD Trichlorofluoromethane C: Standard/Control is not applicable for this analyte
DUPLICATE Vinyl Chloride B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ
STANDARD Volatiles surrogate C: Standard/Control is not applicable for this analyte
DUPLICATE Pet H/C T: Duplicate is not applicable for this analyte

* Indicates NATA accreditation does not cover the performance of this service